



March 29, 2022

Jack Feng  
iHealth Labs, Inc.  
120 San Lucar Ct.  
Sunnyvale, CA 94086

Re: EUA210470/S006  
Trade/Device Name: iHealth COVID-19 Antigen Rapid Test  
Dated: February 11, 2022  
Received: February 11, 2022

Dear Mr. Feng:

This is to notify you that your request to update the iHealth COVID-19 Antigen Rapid Test to extend the shelf-life expiration date of the iHealth COVID-19 Antigen Rapid Test to 9 months, when stored at 2°C – 30°C, based on the results of your ongoing stability studies, is granted. Upon review, we concur that the data and information submitted in EUA210470/S006 support the requested update for the iHealth COVID-19 Antigen Rapid Test. The authorized labeling has been updated to include the 9-month shelf-life. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the iHealth COVID-19 Antigen Rapid Test re-issued on December 22, 2021.

Sincerely yours,

---

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

**iHealth Labs, Inc.: iHealth COVID-19 Antigen Rapid Test  
6-month to 9-month self-life extension granted by the FDA March 29, 2022**

<b>Tests with this printed expiration date (Year-month-day)</b>	<b>Now have this extended expiration date (Year-month-day)</b>
2022-02-20	2022-05-19
2022-02-28	2022-05-27
2022-03-05	2022-06-04
2022-03-08	2022-06-07
2022-05-04	2022-08-03
2022-05-05	2022-08-04
2022-05-07	2022-08-06
2022-05-11	2022-08-10
2022-05-16	2022-08-15
2022-05-17	2022-08-16
2022-05-18	2022-08-17
2022-05-19	2022-08-18
2022-05-20	2022-08-19
2022-05-21	2022-08-20
2022-05-22	2022-08-21
2022-05-23	2022-08-22
2022-05-24	2022-08-23
2022-05-25	2022-08-24
2022-05-26	2022-08-25
2022-05-27	2022-08-26
2022-05-28	2022-08-27
2022-06-02	2022-09-01
2022-06-11	2022-09-10
2022-06-14	2022-09-13
2022-06-15	2022-09-14
2022-06-19	2022-09-18
2022-06-22	2022-09-21
2022-06-23	2022-09-22
2022-06-26	2022-09-25
2022-06-27	2022-09-26
2022-06-28	2022-09-27
2022-06-30	2022-09-29

<b>Tests with this printed expiration date (Year-month-day)</b>	<b>Now have this extended expiration date (Year-month-day)</b>
2022-07-02	2022-10-01
2022-07-04	2022-10-03
2022-07-06	2022-10-05
2022-07-07	2022-10-06
2022-07-09	2022-10-08
2022-07-10	2022-10-09
2022-07-11	2022-10-10
2022-07-12	2022-10-11
2022-07-13	2022-10-12
2022-07-14	2022-10-13
2022-07-16	2022-10-15
2022-07-19	2022-10-18
2022-07-20	2022-10-19
2022-07-21	2022-10-20
2022-07-22	2022-10-21
2022-07-23	2022-10-22
2022-07-24	2022-10-23
2022-07-26	2022-10-25
2022-07-28	2022-10-27
2022-07-29	2022-10-28
2022-07-31	2022-10-30
2022-08-02	2022-11-01
2022-08-03	2022-11-02
2022-08-06	2022-11-05
2022-08-07	2022-11-06
2022-08-09	2022-11-08
2022-08-10	2022-11-09
2022-08-11	2022-11-10
2022-08-13	2022-11-12
2022-08-14	2022-11-13
2022-08-15	2022-11-14
2022-08-16	2022-11-15
2022-08-19	2022-11-18
2022-08-20	2022-11-19
2022-08-21	2022-11-20